

Certificate of Analysis



Product Code: EC-048S
Product: Solus *E. coli* O157 ELISA (1 plate)
LOT / Batch: EE039
Date of Manufacture: 04.04.2022
Expiry Date: 03.10.2023

This batch has been quality control tested to the following criteria:

Batch homogeneity CV%: < 10%.
 A panel of 8 non - *E. coli* O157 *Enterobacteriaceae* spp: Satisfactory
 Non-toxicogenic *E. coli* O157 NCTC 12900: Satisfactory

Performance of In house Reference Standards

Sample	Net Signal OD	Specification OD
F2 <i>E. coli</i> O157 ~10 ⁴ cfu/ml	0.723	>0.300
F2 <i>E. coli</i> O157 ~10 ⁵ cfu/ml	1.972	>0.500
F2 <i>E. coli</i> O157 ~10 ⁶ cfu/ml	2.455	>1.000

Net Optical Densities read at 450nm of samples grown in mTSB + Novobiocin (20mg/L) for 20 hours at 41.5°C

	1	2
A	Negative control 0.006	<i>E5 Serratia liquifaciens</i> 0.024
B	Positive control 2.178	<i>B7 Salmonella</i> Typhimurium 0.016
C	Substrate blank 0.050	<i>A8 E. coli</i> 0.009
D	<i>F2 E. coli</i> O157 NCTC 12900 >2.500	<i>E2 E. coli</i> (Wild Type) 0.023
E	<i>A9 Proteus mirabilis</i> 0.015	Absorbances read at 450nm
F	<i>I3 Citrobacter freundii</i> 0.031	
G	<i>A10 Klebsiella pneumoniae</i> 0.027	
H	<i>E4 Shigella sonnei</i> 0.003	

Certificate of Analysis



Production Batch is approved for release as it gives satisfactory QC performance.

Signed: Technical director/~~Operations~~ Manager/~~Quality~~ Manager

Date:

11/4/22.