

Certificate of Analysis



Product Code: EC-048S
Product: Solus *E. coli* O157 ELISA (1 plate)
LOT / Batch: EE042
Date of Manufacture: 14 NOV 2022
Expiry Date: 13 MAY 2024

This batch has been quality control tested to the following criteria:

Batch homogeneity CV%: < 10%.
A panel of 8 non - *E. coli* O157 *Enterobacteriaceae* spp: Satisfactory
Non-toxicogenic *E. coli* O157 NCTC 12900: Satisfactory

Performance of In house Reference Standards

Sample	Net Signal OD	Specification OD
F2 <i>E. coli</i> O157 ~10 ⁴ cfu/ml	0.544	>0.300
F2 <i>E. coli</i> O157 ~10 ⁵ cfu/ml	1.560	>0.500
F2 <i>E. coli</i> O157 ~10 ⁶ cfu/ml	2.202	>1.000

Net Optical Densities read at 450nm of samples grown in mTSB + Novobiocin (20mg/L) for 20 hours at 41.5°C

	1	2
A	Negative control 0.003	<i>E5 Serratia liquifaciens</i> 0.050
B	Positive control 2.455	<i>B7 Salmonella Typhimurium</i> 0.010
C	Substrate blank 0.051	<i>A8 E. coli</i> 0.000
D	<i>F2 E. coli</i> O157 NCTC 12900 1.899	<i>E2 E. coli</i> (Wild Type) 0.035
E	<i>A9 Proteus mirabilis</i> 0.020	Absorbances read at 450nm
F	<i>I3 Citrobacter freundii</i> 0.037	
G	<i>A10 Klebsiella pneumoniae</i> 0.015	
H	<i>E4 Shigella sonnei</i> 0.005	

Certificate of Analysis



Production Batch is approved for release as it gives satisfactory QC performance.

Signed: Technical director/Operations Manager/Quality Manager

Date: 29/11/22.

