

Certificate of Analysis



Product Code: GA00272010
Product: 3C Standard Solution, 6 x 100ml
Batch Number: 103-23221
Date of Manufacture: 30-MAY-2023
Expiry Date: 29-MAY-2024

This batch has been quality control tested to the following criteria:

3C Value between **10.39** and **10.43**

Test	Result
3C Value	10.4135

Check	Parameter	Result
Dispensing of Product	104g (\pm 1g)	Within Specification
Packing of Product	6 labelled bottles are packed into a labelled box	Conforms

This batch is approved for release, as it gives a satisfactory QC performance.

Signed: ~~Technical Director/Operations Manager/Quality Manager~~

Date: 5th JUNE 2023