

Certificate of Analysis



Product Code: LIS1-0096
Product: Solus ONE *Listeria* (1 plate)
LOT / Batch: NDL032
Date of Manufacture: 03 MAY 2023
Expiry Date: 19 SEP 2024

This batch has been quality control tested to the following criteria:

Batch homogeneity: CV% < 10%.
A panel of 9 non- *Listeria* spp: Satisfactory
A panel of 11 *Listeria* spp: Satisfactory

Performance of In house Reference Standards

Sample	Net Signal OD	OD Specification
<i>D8 L. monocytogenes</i> 4a LMG16783 5x10 ⁴	>2.500	>0.500
<i>D8 L. monocytogenes</i> 4a LMG16783 1x10 ⁵	>2.500	>1.500
<i>D8 L. monocytogenes</i> 4a LMG16783 1x10 ⁶	>2.500	>2.000
<i>G3 L. innocua</i> wild type 5x10 ⁴	1.061	>0.500
<i>G3 L. innocua</i> wild type 1x10 ⁵	1.650	>1.500
<i>G3 L. innocua</i> wild type 1x10 ⁶	>2.500	>2.000
<i>D6 L. welshimeri</i> LMG11389 5x10 ⁴	>2.500	>0.500
<i>D6 L. welshimeri</i> LMG11389 1x10 ⁵	>2.500	>1.400
<i>D6 L. welshimeri</i> LMG11389 1x10 ⁶	>2.500	>2.000

Certificate of Analysis



Net Optical Densities of samples grown in SOLO+ at 30.0°C for 24 hours and boiled at 85-100°C for 20 minutes. Listeria cultures are diluted 1:100 in SOLO+ prior to analysis.

	1	2	3
A	Negative control 0.012	D7 <i>L. monocytogenes</i> >2.500	E8 <i>Brochothrix thermosphacta</i> 0.007
B	Positive control >2.500	D8 <i>L. monocytogenes</i> >2.500	E9 <i>Bacillus cereus</i> 0.005
C	Substrate blank 0.042	D9 <i>L. monocytogenes</i> >2.500	E6 <i>Enterococcus faecalis</i> 0.007
D	G10 <i>L. innocua</i> >2.500	C10 <i>L. monocytogenes</i> >2.500	A8 <i>Escherichia coli</i> 0.006
E	D5 <i>L. ivanovii</i> >2.500	AH10 <i>L. ivanovii</i> <i>londoniensis</i> >2.500	C1 <i>Staphylococcus aureus</i> 0.011
F	D6 <i>L. welshimeri</i> >2.500	G3 <i>L. innocua</i> >2.500	E2 <i>Escherichia coli</i> (Wild Type) 0.012
G	D2 <i>L. grayii</i> >2.500	F1 <i>Streptococcus bovis</i> 0.012	F10 <i>Enterococcus faecalis</i> (Wild type) 0.004
H	D1 <i>L. monocytogenes</i> >2.500	E7 <i>Kurthia zopfii</i> 0.023	Absorbances read at 450nm

Production Batch is approved for release as it gives satisfactory QC performance.

Signed: Technical director/Operations Manager/Quality Manager

D. S. Higgins
Date: 3/7/23.