

Certificate of Analysis



Product Code: SAL-0096S
Product: Solus *Salmonella* ELISA (1 plate)
LOT / Batch: ES200
Date of Manufacture: 27 FEB 2023
Expiry Date: 26 MAY 2024

This batch has been quality control tested to the following criteria:

Batch homogeneity CV%: <10%
A panel of 8 non- *Salmonella* spp: Satisfactory
A panel of 26 *Salmonella* spp: Satisfactory

Performance of In-house Reference Standards

Sample	Net Signal OD	OD Specification
<i>B6 Salmonella</i> Enteritidis 10 ⁶ per ml	>2.500	>2.000
<i>B6 Salmonella</i> Enteritidis 10 ⁷ per ml	>2.500	>2.000
<i>B7 Salmonella</i> Typhimurium 10 ⁶ per ml	>2.500	>1.000
<i>B7 Salmonella</i> Typhimurium 10 ⁷ per ml	>2.500	>2.000
<i>B2 Salmonella</i> Nottingham 10 ⁶ per ml	2.419	>0.350
<i>B2 Salmonella</i> Nottingham 10 ⁷ per ml	>2.500	>1.500

Certificate of Analysis



Net Optical Densities of samples grown in RVS at 41.5°C for 24 hours and boiled at 85-100°C for 20 minutes

	1	2	3	4	5
A	Negative control 0.040	<i>A10 Klebsiella pneumonia</i> 0.028	<i>AH9 Salmonella Nima</i> >2.500	<i>E1 Salmonella Ealing</i> 0.647	<i>AA1 Salmonella Typhimurium</i> (non-motile wild type) >2.500
B	Positive control >2.500	<i>E4 Shigella sonnei</i> 0.025	<i>B8 Salmonella Essen</i> >2.500	<i>B9 Salmonella Kirkee</i> >2.500	<i>A16 Salmonella Hofit</i> >2.500
C	Substrate blank 0.060	<i>E5 Serratia liquifaciens</i> 0.033	<i>AJ9 Salmonella Newport</i> (wild type) >2.500	<i>J8 Salmonella Abaetatuba</i> >2.500	<i>AJ6 Salmonella Menston</i> >2.500
D	<i>A2 Staph. aureus</i> 0.033	<i>B5 Salmonella Heves</i> >2.500	<i>D3 Salmonella Dublin</i> >2.500	<i>J7 Salmonella Bracknell</i> 2.343	<i>V9 Salmonella Minnesota</i> >2.500
E	<i>I3 Citrobacter freundii</i> 0.026	<i>B4 Salmonella Seftenberg</i> >2.500	<i>C5 Salmonella Anatum</i> >2.500	<i>J9 Salmonella Gallinarum</i> (wild type) >2.500	<i>AM9 Salmonella Kentucky</i> >2.500
F	<i>A7 Serratia marcescens</i> 0.025	<i>B3 Salmonella Poona</i> >2.500	<i>AH8 Salmonella Aschersleben</i> 1.367	<i>J10 Salmonella Tranaroa</i> >2.500	Absorbances read at 450nm
G	<i>A8 E. coli</i> 0.024	<i>T10 Salmonella Hadar</i> >2.500	<i>I5 Salmonella arizonae</i> 0.819	<i>K10 Salmonella Virchow</i> >2.500	
H	<i>A9 Proteus mirabilis</i> 0.025	<i>T4 Salmonella Dahlem</i> >2.500	<i>C2 Salmonella Nitra</i> >2.500	<i>T7 Salmonella Dugbe</i> >2.500	

Production Batch is approved for release as it gives satisfactory QC performance.

Signed: Technical Director/Operations Manager/Quality Manager

Date: 11th APRIL 2023