

Certificate of Analysis



Product Code: SAL-0096S
Product: Solus *Salmonella* ELISA (1 plate)
LOT / Batch: ES209-0096
Date of Manufacture: 27 NOV 2023
Expiry Date: 26 FEB 2025

This batch has been quality control tested to the following criteria:

Batch homogeneity CV%: <10%
A panel of 8 non- *Salmonella* spp: Satisfactory
A panel of 26 *Salmonella* spp: Satisfactory

Performance of In-house Reference Standards

Sample	Net Signal OD	OD Specification
B6 <i>Salmonella</i> Enteritidis 10 ⁶ per ml	>2.500	>1.000
B6 <i>Salmonella</i> Enteritidis 10 ⁷ per ml	>2.500	>2.000
B7 <i>Salmonella</i> Typhimurium 10 ⁶ per ml	1.226	>0.750
B7 <i>Salmonella</i> Typhimurium 10 ⁷ per ml	2.427	>2.000
B2 <i>Salmonella</i> Nottingham 10 ⁶ per ml	0.914	>0.350
B2 <i>Salmonella</i> Nottingham 10 ⁷ per ml	2.414	>0.750

Certificate of Analysis



Net Optical Densities of samples grown in RVS at 41.5°C for 24 hours and boiled at 85-100°C for 20 minutes

	1	2	3	4	5
A	Negative control 0.038	A10 <i>Klebsiella pneumoniae</i> 0.040	AH9 <i>Salmonella</i> Nima >2.500	E1 <i>Salmonella</i> Ealing 0.618	AA1 <i>Salmonella</i> Typhimurium (non- motile wild type) >2.500
B	Positive control >2.500	E4 <i>Shigella sonnei</i> 0.040	B8 <i>Salmonella</i> Essen 2.253	B9 <i>Salmonella</i> Kirkee 2.226	A16 <i>Salmonella</i> Hofit 1.429
C	Substrate blank 0.056	E5 <i>Serratia liquifaciens</i> 0.045	AJ9 <i>Salmonella</i> Newport (wild type) 2.117	J8 <i>Salmonella</i> Abaetatuba 2.182	AJ6 <i>Salmonella</i> Menston >2.500
D	A2 <i>Staph. aureus</i> 0.050	B5 <i>Salmonella</i> Heves 2.175	D3 <i>Salmonella</i> Dublin >2.500	J7 <i>Salmonella</i> Bracknell 2.045	V9 <i>Salmonella</i> Minnesota 2.140
E	I3 <i>Citrobacter freundii</i> 0.045	B4 <i>Salmonella</i> Seftenberg >2.500	C5 <i>Salmonella</i> Anatum >2.500	J9 <i>Salmonella</i> Gallinarum (wild type) 2081	AM9 <i>Salmonella</i> Kentucky 2.387
F	A7 <i>Serratia marcescens</i> 0.039	B3 <i>Salmonella</i> Poona >2.500	AH8 <i>Salmonella</i> Aschersleben 1.029	J10 <i>Salmonella</i> Tranaroa 1.926	Absorbances read at 450nm
G	A8 <i>E. coli</i> 0.043	T10 <i>Salmonella</i> Hadar >2.500	I5 <i>Salmonella arizonae</i> 0.449	K10 <i>Salmonella</i> Virchow 2.021	
H	A9 <i>Proteus mirabilis</i> 0.048	T4 <i>Salmonella</i> Dahlem >2.500	C2 <i>Salmonella</i> Nitra >2.500	T7 <i>Salmonella</i> Dugbe >2.500	

Production Batch is approved for release as it gives satisfactory QC performance.

Signed: Technical Director/Operations Manager/Quality Manager

Date:

20/12/23.